

Continuing Calibration Report for SequenceID : FE020725AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE052282.D

Aliphatic C9-C12	5908448.000	60.000	3.679	7.235	98474.133	96184.367	-2.381
Aliphatic C12-C16	4402676.000	40.000	7.236	10.639	110066.900	104714.203	-5.112
Aliphatic C16-C21	6845852.000	60.000	10.640	13.984	114097.533	113022.320	-0.951
Aliphatic C21-C28	8767691.000	80.000	13.985	17.628	109596.138	113998.756	3.862
Aliphatic C28-C40	10777408.000	120.000	17.629	22.712	89811.733	91553.550	1.903
Aliphatic EPH	36702075.000	360.000	3.679	22.712	101950.208	102353.599	0.394