

Continuing Calibration Report for SequenceID : FE020725AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FE052294.D

|                   |              |         |        |        |            |            |        |
|-------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aliphatic C9-C12  | 6005226.000  | 60.000  | 3.679  | 7.235  | 100087.100 | 96184.367  | -4.058 |
| Aliphatic C12-C16 | 4394782.000  | 40.000  | 7.236  | 10.639 | 109869.550 | 104714.203 | -4.923 |
| Aliphatic C16-C21 | 6772395.000  | 60.000  | 10.640 | 13.984 | 112873.250 | 113022.320 | 0.132  |
| Aliphatic C21-C28 | 8555460.000  | 80.000  | 13.985 | 17.628 | 106943.250 | 113998.756 | 6.189  |
| Aliphatic C28-C40 | 10407200.000 | 120.000 | 17.629 | 22.712 | 86726.667  | 91553.550  | 5.272  |
| Aliphatic EPH     | 36135063.000 | 360.000 | 3.679  | 22.712 | 100375.175 | 102353.599 | 1.933  |