

Continuing Calibration Report for SequenceID : FE020725AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE052307.D

Aliphatic C9-C12	6048763.000	60.000	3.679	7.235	100812.717	96184.367	-4.812
Aliphatic C12-C16	4449471.000	40.000	7.236	10.639	111236.775	104714.203	-6.229
Aliphatic C16-C21	6869165.000	60.000	10.640	13.984	114486.083	113022.320	-1.295
Aliphatic C21-C28	8681001.000	80.000	13.985	17.628	108512.513	113998.756	4.813
Aliphatic C28-C40	10518531.000	120.000	17.629	22.712	87654.425	91553.550	4.259
Aliphatic EPH	36566931.000	360.000	3.679	22.712	101574.808	102353.599	0.761