

Continuing Calibration Report for SequenceID : FE021025AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE052310.D

Aliphatic C9-C12	5911603.000	60.000	3.682	7.240	98526.717	96184.367	-2.435
Aliphatic C12-C16	4321034.000	40.000	7.241	10.644	108025.850	104714.203	-3.163
Aliphatic C16-C21	6728632.000	60.000	10.645	13.990	112143.867	113022.320	0.777
Aliphatic C21-C28	8685286.000	80.000	13.991	17.632	108566.075	113998.756	4.766
Aliphatic C28-C40	10801630.000	120.000	17.633	22.724	90013.583	91553.550	1.682
Aliphatic EPH	36448185.000	360.000	3.682	22.724	101244.958	102353.599	1.083