

Continuing Calibration Report for SequenceID : FE021025AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE052316.D

Aliphatic C9-C12	5981300.000	60.000	3.682	7.240	99688.333	96184.367	-3.643
Aliphatic C12-C16	4405264.000	40.000	7.241	10.644	110131.600	104714.203	-5.174
Aliphatic C16-C21	6821789.000	60.000	10.645	13.990	113696.483	113022.320	-0.596
Aliphatic C21-C28	8707040.000	80.000	13.991	17.632	108838.000	113998.756	4.527
Aliphatic C28-C40	10575892.000	120.000	17.633	22.724	88132.433	91553.550	3.737
Aliphatic EPH	36491285.000	360.000	3.682	22.724	101364.681	102353.599	0.966