

Continuing Calibration Report for SequenceID : FF020725AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015484.D

Aromatic C10-C12	5797554.000	40.000	4.505	6.374	144938.850	138122.380	-4.935
Aromatic C12-C16	8485381.000	60.000	6.375	9.068	141423.017	134164.851	-5.410
Aromatic C16-C21	9988045.000	80.000	9.069	13.385	124850.563	123200.544	-1.339
Aromatic C21-C36	21911424.000	180.000	13.386	18.838	121730.133	117776.385	-3.357
Aromatic EPH	46182404.000	360.000	4.505	18.838	128284.456	123973.831	-3.477