

Continuing Calibration Report for SequenceID : FF020725AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015496.D

Aromatic C10-C12	5546906.000	40.000	4.505	6.374	138672.650	138122.380	-0.398
Aromatic C12-C16	8190964.000	60.000	6.375	9.068	136516.067	134164.851	-1.752
Aromatic C16-C21	9921308.000	80.000	9.069	13.385	124016.350	123200.544	-0.662
Aromatic C21-C36	21175365.000	180.000	13.386	18.838	117640.917	117776.385	0.115
Aromatic EPH	44834543.000	360.000	4.505	18.838	124540.397	123973.831	-0.457