

Continuing Calibration Report for SequenceID : FF020725AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015509.D

Aromatic C10-C12	5582246.000	40.000	4.505	6.374	139556.150	138122.380	-1.038
Aromatic C12-C16	8207843.000	60.000	6.375	9.068	136797.383	134164.851	-1.962
Aromatic C16-C21	9858666.000	80.000	9.069	13.385	123233.325	123200.544	-0.027
Aromatic C21-C36	21042540.000	180.000	13.386	18.838	116903.000	117776.385	0.742
Aromatic EPH	44691295.000	360.000	4.505	18.838	124142.486	123973.831	-0.136