

Continuing Calibration Report for SequenceID : FF021025AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015512.D

Aromatic C10-C12	6089611.000	40.000	4.505	6.373	152240.275	138122.380	-10.221
Aromatic C12-C16	8823400.000	60.000	6.374	9.067	147056.667	134164.851	-9.609
Aromatic C16-C21	10470472.000	80.000	9.068	13.385	130880.900	123200.544	-6.234
Aromatic C21-C36	22152034.000	180.000	13.386	18.840	123066.856	117776.385	-4.492
Aromatic EPH	47535517.000	360.000	4.505	18.840	132043.103	123973.831	-6.509