

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** ENTACT
Lab Code: CHEM **Cas No:** Q1308 **SAS No :** Q1308 **SDG No:** Q1308
Sample No : PB166603BS **Datafile:** FD049053.D
Client ID : PB166603BS

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aromatic C10-C12	6.7	4.36	66		(40-140)
Aromatic C12-C16	10	7.56	76		(40-140)
Aromatic C16-C21	16.7	15.2	91		(40-140)
Aromatic C21-C36	26.6	25.1	94		(40-140)