

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLCContract: 68HERH20D0011Lab Code: ACE Case No.: 51979MA No.: _____ SDG No.: MBHHA8Instrument ID: P5Start Date : 02/12/2025Analytical Method: ICP-AESRun Batch : LB134696

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Aluminum	0.999916	0.084538	-0.003535	WLR Blank	Inverse Conc
Antimony	0.999831	0.119089	0.000229	WLR Blank	Inverse Conc
Arsenic	0.999216	249.827480	0.185557	WLR Blank	Inverse Conc
Barium	0.999973	3.102021	0.190906	WLR Blank	Inverse Conc
Beryllium	0.999831	7.143393	0.003033	WLR Blank	Inverse Conc
Boron	0.998377	0.820185	0.001574	WLR Blank	Inverse Conc
Cadmium	0.999979	3.650860	0.000246	WLR Blank	Inverse Conc
Calcium	0.999877	0.062523	0.002956	WLR Blank	Inverse Conc
Chromium	0.999862	0.273930	0.000119	WLR Blank	Inverse Conc
Cobalt	0.999957	0.764477	0.000148	WLR Blank	Inverse Conc
Copper	0.999947	0.176930	0.002823	WLR Blank	Inverse Conc
Iron	0.999779	0.026611	-0.000502	WLR Blank	Inverse Conc
Lead	0.999978	0.235649	0.000341	WLR Blank	Inverse Conc
Lithium	0.999950	0.896910	0.001289	WLR Blank	Inverse Conc
Magnesium	0.999806	0.017426	-0.000467	WLR Blank	Inverse Conc
Manganese	0.999687	0.632409	-0.000302	WLR Blank	Inverse Conc
Molybdenum	0.999955	0.509104	0.000247	WLR Blank	Inverse Conc
Nickel	0.999976	0.691391	-0.000162	WLR Blank	Inverse Conc
Phosphorus	0.999803	0.064574	-0.002219	WLR Blank	Inverse Conc
Potassium	0.999879	0.018201	0.220461	WLR Blank	Inverse Conc
Selenium	0.999905	0.065475	-0.000086	WLR Blank	Inverse Conc
Silicon	0.958286	0.825030	0.012196	WLR Blank	Inverse Conc
Silver	0.999867	0.253554	0.000678	WLR Blank	Inverse Conc
Sodium	0.999891	0.007430	-0.001376	WLR Blank	Inverse Conc
Strontium	0.999911	9.285078	-0.001787	WLR Blank	Inverse Conc
Sulfur	0.999618	0.043815	0.000301	WLR Blank	Inverse Conc
Thallium	0.999623	0.120186	-0.000058	WLR Blank	Inverse Conc
Tin	0.999960	0.171488	0.001581	WLR Blank	Inverse Conc
Vanadium	0.999782	0.190419	0.000573	WLR Blank	Inverse Conc

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51979 MA No.: SDG No.: MBHHA8
Instrument ID: P5 Start Date : 02/12/2025
Analytical Method: ICP-AES Run Batch : LB134696

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Zinc	0.999924	0.152942	-0.000423	WLR Blank	Inverse Conc

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51979 MA No.: SDG No.: MBHHA8
Instrument ID: P5 Start Date : 02/12/2025
Analytical Method: ICP-AES Run Batch : LB134696

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
---------	--------------	-------	-----------	-------------	-----------