

## Initial and Continuing Calibration Verification

**Client:** Tully Environmental, Inc

**SDG No.:** Q1316

**Project:** Transfer Station-SPDES

**RunNo.:** LB134623

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: <b>ICV1</b> Ammonia as N	mg/L	1	1	100	90-110	02/07/2025
Sample ID: <b>CCV1</b> Ammonia as N	mg/L	0.99	1	99	90-110	02/07/2025
Sample ID: <b>CCV2</b> Ammonia as N	mg/L	1	1	100	90-110	02/07/2025
Sample ID: <b>CCV3</b> Ammonia as N	mg/L	1.1	1	110	90-110	02/07/2025
Sample ID: <b>CCV4</b> Ammonia as N	mg/L	0.97	1	97	90-110	02/07/2025