

Matrix Spike Summary

Client:	Tully Environmental, Inc	SDG No.:	Q1316
Project:	Transfer Station-SPDES	Sample ID:	Q1322-01
Client ID:	MANHOLEMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	20.1		0.40	U	20.0	1	101		02/07/2025

Matrix Spike Summary

Client:	Tully Environmental, Inc	SDG No.:	Q1316
Project:	Transfer Station-SPDES	Sample ID:	Q1322-01
Client ID:	MANHOLEMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	20.3		0.40	U	20.0	1	102		02/07/2025

Matrix Spike Summary

Client:	Tully Environmental, Inc	SDG No.:	Q1316
Project:	Transfer Station-SPDES	Sample ID:	Q1325-01
Client ID:	DSN002MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/L	75-125	7.70	OR	6.90	OR	1	1	80		02/07/2025

Matrix Spike Summary

Client:	Tully Environmental, Inc	SDG No.:	Q1316
Project:	Transfer Station-SPDES	Sample ID:	Q1325-01
Client ID:	DSN002MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/L	75-125	8.20	OR	6.90	OR	1	1	130	*	02/07/2025