

Duplicate Sample Summary

Client: Tully Environmental, Inc	SDG No.: Q1316
Project: Transfer Station-SPDES	Sample ID: Q1316-01
Client ID: 001-WILLETS-PT-BLVD(FEB)DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	32.1		31.1		1	3.04		02/06/2025

Duplicate Sample Summary

Client:	Tully Environmental, Inc	SDG No.:	Q1316
Project:	Transfer Station-SPDES	Sample ID:	Q1316-02
Client ID:	002-35TH-AVE(FEB)DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	48.3		47.0		1	2.73		02/10/2025

Duplicate Sample Summary

Client:	Tully Environmental, Inc	SDG No.:	Q1316
Project:	Transfer Station-SPDES	Sample ID:	Q1322-01
Client ID:	MANHOLEMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	20.1		20.3		1	0.99		02/07/2025

Duplicate Sample Summary

Client:	Tully Environmental, Inc	SDG No.:	Q1316
Project:	Transfer Station-SPDES	Sample ID:	Q1325-01
Client ID:	DSN002DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	6.90	OR	6.90	OR	1	0		02/07/2025

Duplicate Sample Summary

Client:	Tully Environmental, Inc	SDG No.:	Q1316
Project:	Transfer Station-SPDES	Sample ID:	Q1325-01
Client ID:	DSN002MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	7.70	OR	8.20	OR	1	6		02/07/2025