

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q1316

Project: Transfer Station-SPDES

Run No.: LB134610

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134610BS							
BOD5	mg/L	198	186		94	1	84.6-115.4	02/06/2025

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q1316

Project: Transfer Station-SPDES

Run No.: LB134620

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134620BS							
Oil and Grease	mg/L	20.0	16.7		84	1	78-114	02/07/2025

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q1316

Project: Transfer Station-SPDES

Run No.: LB134646

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134646BS							
TSS	mg/L	550	532		97	1	90-110	02/10/2025

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q1316

Project: Transfer Station-SPDES

Run No.: LB134623

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB166612BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	02/07/2025