



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Environmental, Inc **SDG No.:** Q1316
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1316 **SAS No.:** Q1316
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	20.3	20.0	102	40 - 160	P	02/07/2025	12:35	LB134636
	Cadmium	5.87	6.0	98	40 - 160	P	02/07/2025	12:35	LB134636
	Chromium	9.48	10.0	95	40 - 160	P	02/07/2025	12:35	LB134636
	Copper	21.0	20.0	105	40 - 160	P	02/07/2025	12:35	LB134636
	Iron	89.1	100	89	40 - 160	P	02/07/2025	12:35	LB134636
	Lead	10.9	12.0	91	40 - 160	P	02/07/2025	12:35	LB134636
	Selenium	19.8	20.0	99	40 - 160	P	02/07/2025	12:35	LB134636