



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q1316

Contract: TULL01 Lab Code: CHEM

Case No.: Q1316

SAS No.: Q1316

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	02/07/2025	12:31	LB134636
	Cadmium	6.00	+/-6.00	U	6.00	P	02/07/2025	12:31	LB134636
	Chromium	10.0	+/-10.0	U	10.0	P	02/07/2025	12:31	LB134636
	Copper	20.0	+/-20.0	U	20.0	P	02/07/2025	12:31	LB134636
	Iron	100	+/-100	U	100	P	02/07/2025	12:31	LB134636
	Lead	12.0	+/-12.0	U	12.0	P	02/07/2025	12:31	LB134636
	Selenium	20.0	+/-20.0	U	20.0	P	02/07/2025	12:31	LB134636

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc		SDG No.: Q1316							
Contract: TULL01	Lab Code: CHEM	Case No.: Q1316	SAS No.: Q1316						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	02/07/2025	13:01	LB134636
	Cadmium	6.00	+/-6.00	U	6.00	P	02/07/2025	13:01	LB134636
	Chromium	10.0	+/-10.0	U	10.0	P	02/07/2025	13:01	LB134636
	Copper	20.0	+/-20.0	U	20.0	P	02/07/2025	13:01	LB134636
	Iron	100	+/-100	U	100	P	02/07/2025	13:01	LB134636
	Lead	12.0	+/-12.0	U	12.0	P	02/07/2025	13:01	LB134636
	Selenium	20.0	+/-20.0	U	20.0	P	02/07/2025	13:01	LB134636
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	02/07/2025	13:53	LB134636
	Cadmium	0.41	+/-6.00	J	6.00	P	02/07/2025	13:53	LB134636
	Chromium	10.0	+/-10.0	U	10.0	P	02/07/2025	13:53	LB134636
	Copper	20.0	+/-20.0	U	20.0	P	02/07/2025	13:53	LB134636
	Iron	100	+/-100	U	100	P	02/07/2025	13:53	LB134636
	Lead	12.0	+/-12.0	U	12.0	P	02/07/2025	13:53	LB134636
	Selenium	20.0	+/-20.0	U	20.0	P	02/07/2025	13:53	LB134636
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	02/07/2025	14:45	LB134636
	Cadmium	0.42	+/-6.00	J	6.00	P	02/07/2025	14:45	LB134636
	Chromium	10.0	+/-10.0	U	10.0	P	02/07/2025	14:45	LB134636
	Copper	20.0	+/-20.0	U	20.0	P	02/07/2025	14:45	LB134636
	Iron	100	+/-100	U	100	P	02/07/2025	14:45	LB134636
	Lead	12.0	+/-12.0	U	12.0	P	02/07/2025	14:45	LB134636
	Selenium	20.0	+/-20.0	U	20.0	P	02/07/2025	14:45	LB134636
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	02/07/2025	15:37	LB134636
	Cadmium	0.26	+/-6.00	J	6.00	P	02/07/2025	15:37	LB134636
	Chromium	10.0	+/-10.0	U	10.0	P	02/07/2025	15:37	LB134636
	Copper	20.0	+/-20.0	U	20.0	P	02/07/2025	15:37	LB134636
	Iron	100	+/-100	U	100	P	02/07/2025	15:37	LB134636
	Lead	12.0	+/-12.0	U	12.0	P	02/07/2025	15:37	LB134636
	Selenium	20.0	+/-20.0	U	20.0	P	02/07/2025	15:37	LB134636
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	02/07/2025	16:28	LB134636
	Cadmium	6.00	+/-6.00	U	6.00	P	02/07/2025	16:28	LB134636
	Chromium	10.0	+/-10.0	U	10.0	P	02/07/2025	16:28	LB134636
	Copper	20.0	+/-20.0	U	20.0	P	02/07/2025	16:28	LB134636
	Iron	100	+/-100	U	100	P	02/07/2025	16:28	LB134636
	Lead	12.0	+/-12.0	U	12.0	P	02/07/2025	16:28	LB134636
	Selenium	20.0	+/-20.0	U	20.0	P	02/07/2025	16:28	LB134636
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	02/07/2025	17:20	LB134636
	Cadmium	0.28	+/-6.00	J	6.00	P	02/07/2025	17:20	LB134636
	Chromium	10.0	+/-10.0	U	10.0	P	02/07/2025	17:20	LB134636
	Copper	20.0	+/-20.0	U	20.0	P	02/07/2025	17:20	LB134636

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc		SDG No.: Q1316							
Contract: TULL01	Lab Code: CHEM	Case No.: Q1316	SAS No.: Q1316						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Iron	100	+/-100	U	100	P	02/07/2025	17:20	LB134636
	Lead	12.0	+/-12.0	U	12.0	P	02/07/2025	17:20	LB134636
	Selenium	20.0	+/-20.0	U	20.0	P	02/07/2025	17:20	LB134636
CCB07	Arsenic	20.0	+/-20.0	U	20.0	P	02/07/2025	18:10	LB134636
	Cadmium	6.00	+/-6.00	U	6.00	P	02/07/2025	18:10	LB134636
	Chromium	10.0	+/-10.0	U	10.0	P	02/07/2025	18:10	LB134636
	Copper	20.0	+/-20.0	U	20.0	P	02/07/2025	18:10	LB134636
	Iron	100	+/-100	U	100	P	02/07/2025	18:10	LB134636
	Lead	12.0	+/-12.0	U	12.0	P	02/07/2025	18:10	LB134636
	Selenium	20.0	+/-20.0	U	20.0	P	02/07/2025	18:10	LB134636
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	02/07/2025	18:27	LB134636
	Cadmium	0.27	+/-6.00	J	6.00	P	02/07/2025	18:27	LB134636
	Chromium	10.0	+/-10.0	U	10.0	P	02/07/2025	18:27	LB134636
	Copper	20.0	+/-20.0	U	20.0	P	02/07/2025	18:27	LB134636
	Iron	100	+/-100	U	100	P	02/07/2025	18:27	LB134636
	Lead	12.0	+/-12.0	U	12.0	P	02/07/2025	18:27	LB134636
	Selenium	20.0	+/-20.0	U	20.0	P	02/07/2025	18:27	LB134636