

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc **SDG No.:** Q1316
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1316 **SAS No.:** Q1316
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	2.16			-20	20	02/07/2025	12:40	LB134636
	Cadmium	2.97	1.0	297	-5	7	02/07/2025	12:40	LB134636
	Chromium	58.5	52.0	112	42	62	02/07/2025	12:40	LB134636
	Copper	9.83	2.0	492	-18	22	02/07/2025	12:40	LB134636
	Iron	101000	101000	100	85600	116500	02/07/2025	12:40	LB134636
	Lead	8.02			-12	12	02/07/2025	12:40	LB134636
	Selenium	-11.1			-20	20	02/07/2025	12:40	LB134636
ICSAB01	Arsenic	108	104	104	88.4	120	02/07/2025	12:44	LB134636
	Cadmium	1020	972	105	826	1120	02/07/2025	12:44	LB134636
	Chromium	575	542	106	460	624	02/07/2025	12:44	LB134636
	Copper	519	511	102	434	588	02/07/2025	12:44	LB134636
	Iron	98800	99300	100	84400	114500	02/07/2025	12:44	LB134636
	Lead	56.3	49.0	115	37	61	02/07/2025	12:44	LB134636
	Selenium	41.6	46.0	90	26	66	02/07/2025	12:44	LB134636
ICSA	Arsenic	2.98			-20	20	02/07/2025	12:48	LB134636
	Cadmium	4.47	1.0	447	-5	7	02/07/2025	12:48	LB134636
	Chromium	51.0	52.0	98	42	62	02/07/2025	12:48	LB134636
	Copper	-15.3	2.0	765	-18	22	02/07/2025	12:48	LB134636
	Iron	99100	101000	98	85600	116500	02/07/2025	12:48	LB134636
	Lead	-20.9			-12	12	02/07/2025	12:48	LB134636
	Selenium	25.1			-20	20	02/07/2025	12:48	LB134636
ICSAB	Arsenic	117	104	112	88.4	120	02/07/2025	12:52	LB134636
	Cadmium	1040	972	107	826	1120	02/07/2025	12:52	LB134636
	Chromium	586	542	108	460	624	02/07/2025	12:52	LB134636
	Copper	526	511	103	434	588	02/07/2025	12:52	LB134636
	Iron	98400	99300	99	84400	114500	02/07/2025	12:52	LB134636
	Lead	30.4	49.0	62	37	61	02/07/2025	12:52	LB134636
	Selenium	49.2	46.0	107	26	66	02/07/2025	12:52	LB134636