

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Tully Environmental, Inc **level:** low **sdg no.:** Q1316
contract: TULL01 **lab code:** CHEM **case no.:** Q1316 **sas no.:** Q1316
matrix: Water **sample id:** Q1316-02 **client id:** 002-35TH-AVE(FEB)MS
Percent Solids for Sample: NA **Spiked ID:** Q1316-02MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Arsenic	ug/L	75 - 125	435		7.45	J	400	107		P
Cadmium	ug/L	75 - 125	99.0		0.26	J	100	99		P
Chromium	ug/L	75 - 125	198		3.15	J	200	98		P
Copper	ug/L	75 - 125	155		8.82	J	150	97		P
Iron	ug/L	75 - 125	6690		5130		1500	104		P
Lead	ug/L	75 - 125	491		12.6		500	96		P
Selenium	ug/L	75 - 125	1000		10.0	U	1000	100		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Tully Environmental, Inc **level:** low **sdg no.:** Q1316
contract: TULL01 **lab code:** CHEM **case no.:** Q1316 **sas no.:** Q1316
matrix: Water **sample id:** Q1316-02 **client id:** 002-35TH-AVE(FEB)MSD
Percent Solids for Sample: NA **Spiked ID:** Q1316-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Arsenic	ug/L	75 - 125	435		7.45	J	400	107		P
Cadmium	ug/L	75 - 125	99.0		0.26	J	100	99		P
Chromium	ug/L	75 - 125	199		3.15	J	200	98		P
Copper	ug/L	75 - 125	155		8.82	J	150	98		P
Iron	ug/L	75 - 125	6410		5130		1500	85		P
Lead	ug/L	75 - 125	492		12.6		500	96		P
Selenium	ug/L	75 - 125	1000		10.0	U	1000	100		P