

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Environmental, Inc **SDG No.:** Q1316
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1316 **SAS No.:** Q1316

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166597BS							
Arsenic	ug/L	400	387		97	85 - 115	P
Cadmium	ug/L	100	94.3		94	85 - 115	P
Chromium	ug/L	200	200		100	85 - 115	P
Copper	ug/L	150	152		101	85 - 115	P
Iron	ug/L	1500	1390		93	85 - 115	P
Lead	ug/L	500	469		94	85 - 115	P
Selenium	ug/L	1000	978		98	85 - 115	P