

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX011625

Review By	Mahesh Dadoda	Review On	1/17/2025 6:39:05 AM
Supervise By	Semsettin Yesilyurt	Supervise On	1/17/2025 6:41:46 AM
SubDirectory	VX011625	HP Acquire Method	HP Processing Method 624X011625W.M

STD. NAME	STD REF.#
Tune/Reschk	VP132565
Initial Calibration Stds	VP132567,VP132568,VP132569,VP132570,VP132571
CCC	VP132566,VP132573
Internal Standard/PEM	
ICV/I.BLK	VP132572
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX044658.D	16 Jan 2025 08:02		 	Ok
2	VSTDIC005	VSTDIC005	VX044659.D	16 Jan 2025 08:39	V13516	 	Ok,M
3	VSTDIC020	VSTDIC020	VX044660.D	16 Jan 2025 09:02		 	Ok,M
4	VSTDIC050	VSTDIC050	VX044661.D	16 Jan 2025 09:25		 	Ok,M
5	VSTDIC100	VSTDIC100	VX044662.D	16 Jan 2025 09:49		 	Ok,M
6	VSTDIC150	VSTDIC150	VX044663.D	16 Jan 2025 10:12		 	Ok,M
7	IBLK	IBLK	VX044664.D	16 Jan 2025 10:35		 	Ok
8	VSTDICV020	ICVVX011625	VX044665.D	16 Jan 2025 11:45		 	Ok,M
9	VX0116WBS01	VX0116WBS01	VX044666.D	16 Jan 2025 12:15		 	Ok,M
10	VX0116WBSD01	VX0116WBSD01	VX044667.D	16 Jan 2025 12:42		 	Ok,M
11	VX0116WBL01	VX0116WBL01	VX044668.D	16 Jan 2025 13:05		 	Ok
12	Q1083-01	001-WILLETS-PT-BLVD	VX044669.D	16 Jan 2025 13:33	vial A pH<2	 	Ok
13	Q1083-02	002-35TH-AVE(DEC)	VX044670.D	16 Jan 2025 13:55	vial A pH<2	 	Ok
14	IBLK	IBLK	VX044671.D	16 Jan 2025 14:18		 	Ok
15	P4368-09 2.5PPB	MDL-WATER-03-QT4-2	VX044672.D	16 Jan 2025 14:46		 	Ok,M
16	VSTDIC020	VSTDIC020EC	VX044673.D	16 Jan 2025 16:04		 	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX020725

Review By	John Carlone	Review On	2/7/2025 1:17:07 PM
Supervise By		Supervise On	
SubDirectory	VX020725	HP Acquire Method	HP Processing Method 624X011625W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP132924
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP132925,VP132926

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX044855.D	07 Feb 2025 09:33		JC/MD	Ok
2	VSTDCCC020	VSTDCCC020	VX044856.D	07 Feb 2025 10:01	V13516	JC/MD	Ok,NR
3	VX0207WBS01	VX0207WBS01	VX044857.D	07 Feb 2025 10:33		JC/MD	Ok,NR
4	VX0207WBSD01	VX0207WBSD01	VX044858.D	07 Feb 2025 11:01		JC/MD	Ok,NR
5	VX0207WBL01	VX0207WBL01	VX044859.D	07 Feb 2025 11:24		JC/MD	Ok
6	Q1319-01	001-WILLETS-PT-BLVD	VX044860.D	07 Feb 2025 11:53	vial A pH<2	JC/MD	Ok,NS
7	Q1319-02	002-35TH-AVE(JAN)	VX044861.D	07 Feb 2025 12:16	vial A pH<2	JC/MD	Ok
8	Q1316-01	001-WILLETS-PT-BLVD	VX044862.D	07 Feb 2025 12:39		JC/MD	Ok
9	Q1316-02	002-35TH-AVE(FEB)	VX044863.D	07 Feb 2025 13:02		JC/MD	Ok
10	IBLK	IBLK	VX044864.D	07 Feb 2025 13:25		JC/MD	Ok
11	Q1322-01	MANHOLE	VX044865.D	07 Feb 2025 13:48		JC/MD	Ok,NR
12	VSTDCCC020	VSTDCCC020EC	VX044866.D	07 Feb 2025 15:43		JC/MD	Ok,NR

M : Manual Integration