

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1319

Client: Tully Environmental, Inc

Analytical Method: SW624.1

Datafile : VX044857.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0207WBS01	Benzene	20	21.0	ug/L	105			65	135	
	Toluene	20	20.7	ug/L	104			70	130	
	Ethyl Benzene	20	20.8	ug/L	104			60	140	
	m/p-Xylenes	40	43.0	ug/L	108			87	111	
	o-Xylene	20	21.5	ug/L	108			87	111	