

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

**SDG No.:** Q1319

**Client:** Tully Environmental, Inc

**Analytical Method:** SW624.1

**Datafile :** VX044858.D

| Lab Sample ID | Parameter     | Spike | Result | Unit | Rec | RPD | Qual | Low | Limits<br>High | RPD |
|---------------|---------------|-------|--------|------|-----|-----|------|-----|----------------|-----|
| VX0207WBSD01  | Benzene       | 20    | 21.6   | ug/L | 108 | 3   |      | 65  | 135            | 20  |
|               | Toluene       | 20    | 20.7   | ug/L | 104 | 0   |      | 70  | 130            | 20  |
|               | Ethyl Benzene | 20    | 20.8   | ug/L | 104 | 0   |      | 60  | 140            | 20  |
|               | m/p-Xylenes   | 40    | 42.8   | ug/L | 107 | 1   |      | 87  | 111            | 20  |
|               | o-Xylene      | 20    | 21.2   | ug/L | 106 | 2   |      | 87  | 111            | 20  |