

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL01
 Lab Code: CHEM Case No.: Q1319 SAS No.: Q1319 SDG NO.: Q1319
 Lab File ID: VX044856.D Date Analyzed: 02/07/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:01
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	21982	4.89	127777	6.75	110326	10.05
UPPER LIMIT	43964	5.391	255554	7.251	220652	10.549
LOWER LIMIT	10991	4.391	63888.5	6.251	55163	9.549
EPA SAMPLE NO.						
001-WILLETS-PT-BLVD (JAN)	23812	4.89	133041	6.76	120992	10.05
002-35TH-AVE (JAN)	17858	4.90	100352	6.76	88457	10.05
VX0207WBL01	20510	4.90	114099	6.76	99648	10.05
VX0207WBS01	21269	4.89	121245	6.75	109254	10.05
VX0207WBSD01	19387	4.89	108776	6.76	100274	10.05

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.