

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TULL01  
Lab Code: CHEM Case No.: Q1319 SAS No.: Q1319 SDG NO.: Q1319  
Lab File ID: VX044658.D BFB Injection Date: 01/16/2025  
Instrument ID: MSVOA\_X BFB Injection Time: 08:02  
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 18.2                 |
| 75  | 30.0 - 60.0% of mass 95            | 48.5                 |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 7.2                  |
| 173 | Less than 2.0% of mass 174         | 0.5 ( 0.7 ) 1        |
| 174 | 50.0 - 100.0% of mass 95           | 66.7                 |
| 175 | 5.0 - 9.0% of mass 174             | 4.7 ( 7 ) 1          |
| 176 | 95.0 - 101.0% of mass 174          | 64.7 ( 97 ) 1        |
| 177 | 5.0 - 9.0% of mass 176             | 4.3 ( 6.6 ) 2        |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| EPA<br>SAMPLE NO. | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED | TIME<br>ANALYZED |
|-------------------|------------------|----------------|------------------|------------------|
| VSTDIC005         | VSTDIC005        | VX044659.D     | 01/16/2025       | 08:39            |
| VSTDIC020         | VSTDIC020        | VX044660.D     | 01/16/2025       | 09:02            |
| VSTDIC050         | VSTDIC050        | VX044661.D     | 01/16/2025       | 09:25            |
| VSTDIC100         | VSTDIC100        | VX044662.D     | 01/16/2025       | 09:49            |
| VSTDIC150         | VSTDIC150        | VX044663.D     | 01/16/2025       | 10:12            |