



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Spectra East Inc. **SDG No.:** Q1322
Contract: SPEC01 **Lab Code:** CHEM **Case No.:** Q1322 **SAS No.:** Q1322
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.055	0.05	111	40 - 160	CV	02/11/2025	15:28	LB134671
CRI01	Arsenic	22.9	20.0	114	40 - 160	P	02/18/2025	12:18	LB134743
	Beryllium	5.61	6.0	94	40 - 160	P	02/18/2025	12:18	LB134743
	Cadmium	6.11	6.0	102	40 - 160	P	02/18/2025	12:18	LB134743
	Chromium	10.3	10.0	103	40 - 160	P	02/18/2025	12:18	LB134743
	Copper	21.9	20.0	110	40 - 160	P	02/18/2025	12:18	LB134743
	Lead	13.4	12.0	111	40 - 160	P	02/18/2025	12:18	LB134743
	Nickel	39.0	40.0	98	40 - 160	P	02/18/2025	12:18	LB134743
	Selenium	19.6	20.0	98	40 - 160	P	02/18/2025	12:18	LB134743
	Silver	10.3	10.0	103	40 - 160	P	02/18/2025	12:18	LB134743
	Zinc	44.3	40.0	111	40 - 160	P	02/18/2025	12:18	LB134743