



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Spectra East Inc.</u>				SDG No.:	<u>Q1322</u>			
Contract:	<u>SPEC01</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>Q1322</u>	SAS No.:	<u>Q1322</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB96	Mercury	0.20	+/-0.20	U	0.20	CV	02/11/2025	15:19	LB134671

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Spectra East Inc.

SDG No.: Q1322

Contract: SPEC01

Lab Code: CHEM

Case No.: Q1322

SAS No.: Q1322

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB56	Mercury	0.20	+/-0.20	U	0.20	CV	02/11/2025	15:24	LB134671
CCB57	Mercury	0.20	+/-0.20	U	0.20	CV	02/11/2025	16:02	LB134671
CCB58	Mercury	0.20	+/-0.20	U	0.20	CV	02/11/2025	16:22	LB134671

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Spectra East Inc.</u>		SDG No.: <u>Q1322</u>							
Contract: <u>SPEC01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1322</u>	SAS No.: <u>Q1322</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	02/18/2025	12:14	LB134743
	Beryllium	6.00	+/-6.00	U	6.00	P	02/18/2025	12:14	LB134743
	Cadmium	6.00	+/-6.00	U	6.00	P	02/18/2025	12:14	LB134743
	Chromium	10.0	+/-10.0	U	10.0	P	02/18/2025	12:14	LB134743
	Copper	20.0	+/-20.0	U	20.0	P	02/18/2025	12:14	LB134743
	Lead	12.0	+/-12.0	U	12.0	P	02/18/2025	12:14	LB134743
	Nickel	40.0	+/-40.0	U	40.0	P	02/18/2025	12:14	LB134743
	Selenium	20.0	+/-20.0	U	20.0	P	02/18/2025	12:14	LB134743
	Silver	10.0	+/-10.0	U	10.0	P	02/18/2025	12:14	LB134743
	Zinc	40.0	+/-40.0	U	40.0	P	02/18/2025	12:14	LB134743

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Spectra East Inc.</u>		SDG No.: <u>Q1322</u>							
Contract: <u>SPEC01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1322</u>	SAS No.: <u>Q1322</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	02/18/2025	12:45	LB134743
	Beryllium	6.00	+/-6.00	U	6.00	P	02/18/2025	12:45	LB134743
	Cadmium	6.00	+/-6.00	U	6.00	P	02/18/2025	12:45	LB134743
	Chromium	10.0	+/-10.0	U	10.0	P	02/18/2025	12:45	LB134743
	Copper	20.0	+/-20.0	U	20.0	P	02/18/2025	12:45	LB134743
	Lead	12.0	+/-12.0	U	12.0	P	02/18/2025	12:45	LB134743
	Nickel	40.0	+/-40.0	U	40.0	P	02/18/2025	12:45	LB134743
	Selenium	20.0	+/-20.0	U	20.0	P	02/18/2025	12:45	LB134743
	Silver	10.0	+/-10.0	U	10.0	P	02/18/2025	12:45	LB134743
	Zinc	40.0	+/-40.0	U	40.0	P	02/18/2025	12:45	LB134743
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	02/18/2025	13:36	LB134743
	Beryllium	6.00	+/-6.00	U	6.00	P	02/18/2025	13:36	LB134743
	Cadmium	6.00	+/-6.00	U	6.00	P	02/18/2025	13:36	LB134743
	Chromium	10.0	+/-10.0	U	10.0	P	02/18/2025	13:36	LB134743
	Copper	20.0	+/-20.0	U	20.0	P	02/18/2025	13:36	LB134743
	Lead	12.0	+/-12.0	U	12.0	P	02/18/2025	13:36	LB134743
	Nickel	40.0	+/-40.0	U	40.0	P	02/18/2025	13:36	LB134743
	Selenium	20.0	+/-20.0	U	20.0	P	02/18/2025	13:36	LB134743
	Silver	10.0	+/-10.0	U	10.0	P	02/18/2025	13:36	LB134743
	Zinc	40.0	+/-40.0	U	40.0	P	02/18/2025	13:36	LB134743
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	02/18/2025	14:27	LB134743
	Beryllium	6.00	+/-6.00	U	6.00	P	02/18/2025	14:27	LB134743
	Cadmium	6.00	+/-6.00	U	6.00	P	02/18/2025	14:27	LB134743
	Chromium	10.0	+/-10.0	U	10.0	P	02/18/2025	14:27	LB134743
	Copper	20.0	+/-20.0	U	20.0	P	02/18/2025	14:27	LB134743
	Lead	12.0	+/-12.0	U	12.0	P	02/18/2025	14:27	LB134743
	Nickel	40.0	+/-40.0	U	40.0	P	02/18/2025	14:27	LB134743
	Selenium	20.0	+/-20.0	U	20.0	P	02/18/2025	14:27	LB134743
	Silver	10.0	+/-10.0	U	10.0	P	02/18/2025	14:27	LB134743
	Zinc	40.0	+/-40.0	U	40.0	P	02/18/2025	14:27	LB134743
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	02/18/2025	15:23	LB134743
	Beryllium	6.00	+/-6.00	U	6.00	P	02/18/2025	15:23	LB134743
	Cadmium	6.00	+/-6.00	U	6.00	P	02/18/2025	15:23	LB134743
	Chromium	10.0	+/-10.0	U	10.0	P	02/18/2025	15:23	LB134743
	Copper	20.0	+/-20.0	U	20.0	P	02/18/2025	15:23	LB134743
	Lead	12.0	+/-12.0	U	12.0	P	02/18/2025	15:23	LB134743
	Nickel	40.0	+/-40.0	U	40.0	P	02/18/2025	15:23	LB134743
	Selenium	20.0	+/-20.0	U	20.0	P	02/18/2025	15:23	LB134743
	Silver	10.0	+/-10.0	U	10.0	P	02/18/2025	15:23	LB134743

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Spectra East Inc.</u>		SDG No.: <u>Q1322</u>							
Contract: <u>SPEC01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1322</u>	SAS No.: <u>Q1322</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Zinc	40.0	+/-40.0	U	40.0	P	02/18/2025	15:23	LB134743
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	02/18/2025	16:11	LB134743
	Beryllium	6.00	+/-6.00	U	6.00	P	02/18/2025	16:11	LB134743
	Cadmium	6.00	+/-6.00	U	6.00	P	02/18/2025	16:11	LB134743
	Chromium	10.0	+/-10.0	U	10.0	P	02/18/2025	16:11	LB134743
	Copper	20.0	+/-20.0	U	20.0	P	02/18/2025	16:11	LB134743
	Lead	12.0	+/-12.0	U	12.0	P	02/18/2025	16:11	LB134743
	Nickel	40.0	+/-40.0	U	40.0	P	02/18/2025	16:11	LB134743
	Selenium	20.0	+/-20.0	U	20.0	P	02/18/2025	16:11	LB134743
	Silver	10.0	+/-10.0	U	10.0	P	02/18/2025	16:11	LB134743
	Zinc	40.0	+/-40.0	U	40.0	P	02/18/2025	16:11	LB134743

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Spectra East Inc.

SDG No.: Q1322

Contract: SPEC01

Lab Code: CHEM

Case No.: Q1322

SAS No.: Q1322

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------