

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Spectra East Inc.

SDG No.: Q1322

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166684BL		WATER		Batch Number:	PB166684		Prep Date:	02/11/2025	
	Mercury	0.20	<0.20	U	0.20	CV	02/11/2025	15:35	LB134671

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Spectra East Inc.

SDG No.: Q1322

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166737BL	WATER			Batch Number:	PB166737		Prep Date:	02/17/2025	
	Arsenic	10.0	<10.0	U	10.0	P	02/18/2025	13:15	LB134743
	Beryllium	3.00	<3.00	U	3.00	P	02/18/2025	13:15	LB134743
	Cadmium	3.00	<3.00	U	3.00	P	02/18/2025	13:15	LB134743
	Chromium	0.52	<5.00	J	5.00	P	02/18/2025	13:15	LB134743
	Copper	10.0	<10.0	U	10.0	P	02/18/2025	13:15	LB134743
	Lead	6.00	<6.00	U	6.00	P	02/18/2025	13:15	LB134743
	Nickel	20.0	<20.0	U	20.0	P	02/18/2025	13:15	LB134743
	Selenium	10.0	<10.0	U	10.0	P	02/18/2025	13:15	LB134743
	Silver	5.00	<5.00	U	5.00	P	02/18/2025	13:15	LB134743
	Zinc	1.52	<20.0	J	20.0	P	02/18/2025	13:15	LB134743