

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Spectra East Inc. **Level:** LOW **SDG No.:** Q1322
Contract: SPEC01 **Lab Code:** CHEM **Case No.:** Q1322 **SAS No.:** Q1322
Matrix: Water **Sample ID:** Q1322-01 **Client ID:** MANHOLEDUP
Percent Solids for Sample: NA **Duplicate ID** Q1322-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	10.0	U	10.0	U			P
Beryllium	ug/L	20	3.00	U	3.00	U			P
Cadmium	ug/L	20	3.00	U	3.00	U			P
Chromium	ug/L	20	5.00	U	5.00	U			P
Copper	ug/L	20	56.3		55.9		1		P
Lead	ug/L	20	6.00	U	6.00	U			P
Mercury	ug/L	20	0.20	U	0.20	U			CV
Nickel	ug/L	20	20.0	U	20.0	U			P
Selenium	ug/L	20	10.0	U	10.0	U			P
Silver	ug/L	20	5.00	U	5.00	U			P
Zinc	ug/L	20	140		142		1		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Spectra East Inc. **Level:** LOW **SDG No.:** Q1322
Contract: SPEC01 **Lab Code:** CHEM **Case No.:** Q1322 **SAS No.:** Q1322
Matrix: Water **Sample ID:** Q1322-01MS **Client ID:** MANHOLEMSD
Percent Solids for Sample: NA **Duplicate ID** Q1322-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	390		395		1		P
Beryllium	ug/L	20	88.1		88.5		0		P
Cadmium	ug/L	20	92.8		93.1		0		P
Chromium	ug/L	20	187		190		2		P
Copper	ug/L	20	192		194		1		P
Lead	ug/L	20	462		463		0		P
Mercury	ug/L	20	3.55		3.79		7		CV
Nickel	ug/L	20	236		236		0		P
Selenium	ug/L	20	917		936		2		P
Silver	ug/L	20	35.0		35.9		3		P
Zinc	ug/L	20	223		230		3		P