

CALIBRATION VERIFICATION SUMMARY

Contract: SPEC01

Lab Code: CHEM Case No.: Q1322 SAS No.: Q1322 SDG NO.: Q1322

Continuing Calib Date: 02/07/2025 Initial Calibration Date(s): 01/28/2025 01/28/2025

Continuing Calib Time: 10:02 Initial Calibration Time(s): 09:37 17:30

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.69	5.69	5.59	5.79	0.00
Aroclor-1248-2 (2)	5.96	5.97	5.87	6.07	0.01
Aroclor-1248-3 (3)	6.16	6.17	6.07	6.27	0.01
Aroclor-1248-4 (4)	6.56	6.57	6.47	6.67	0.01
Aroclor-1248-5 (5)	6.60	6.61	6.51	6.71	0.01
Tetrachloro-m-xylene	4.53	4.54	4.44	4.64	0.01
Decachlorobiphenyl	10.27	10.28	10.18	10.38	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: SPEC01

Lab Code: CHEM Case No.: Q1322 SAS No.: Q1322 SDG NO.: Q1322

Continuing Calib Date: 02/07/2025 Initial Calibration Date(s): 01/28/2025 01/28/2025

Continuing Calib Time: 10:02 Initial Calibration Time(s): 09:37 17:30

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.93	4.93	4.83	5.03	0.00
Aroclor-1248-2	(2)	5.17	5.17	5.07	5.27	0.00
Aroclor-1248-3	(3)	5.21	5.22	5.12	5.32	0.01
Aroclor-1248-4	(4)	5.38	5.39	5.29	5.49	0.01
Aroclor-1248-5	(5)	5.78	5.78	5.68	5.88	0.00
Tetrachloro-m-xylene		3.84	3.84	3.74	3.94	0.00
Decachlorobiphenyl		8.90	8.91	8.81	9.01	0.01