

### CALIBRATION VERIFICATION SUMMARY

Contract: SPEC01

Lab Code: CHEM Case No.: Q1322 SAS No.: Q1322 SDG NO.: Q1322

Continuing Calib Date: 02/07/2025 Initial Calibration Date(s): 01/28/2025 01/28/2025

Continuing Calib Time: 14:08 Initial Calibration Time(s): 09:37 17:30

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1248-1       | (1) | 5.69       | 5.69      | 5.59      | 5.79  | 0.00       |
| Aroclor-1248-2       | (2) | 5.96       | 5.97      | 5.87      | 6.07  | 0.01       |
| Aroclor-1248-3       | (3) | 6.16       | 6.17      | 6.07      | 6.27  | 0.01       |
| Aroclor-1248-4       | (4) | 6.56       | 6.57      | 6.47      | 6.67  | 0.01       |
| Aroclor-1248-5       | (5) | 6.60       | 6.61      | 6.51      | 6.71  | 0.01       |
| Tetrachloro-m-xylene |     | 4.54       | 4.54      | 4.44      | 4.64  | 0.01       |
| Decachlorobiphenyl   |     | 10.27      | 10.28     | 10.18     | 10.38 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Contract: SPEC01

Lab Code: CHEM Case No.: Q1322 SAS No.: Q1322 SDG NO.: Q1322

Continuing Calib Date: 02/07/2025 Initial Calibration Date(s): 01/28/2025 01/28/2025

Continuing Calib Time: 14:08 Initial Calibration Time(s): 09:37 17:30

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1248-1       | (1) | 4.93       | 4.93      | 4.83      | 5.03 | 0.00       |
| Aroclor-1248-2       | (2) | 5.17       | 5.17      | 5.07      | 5.27 | 0.00       |
| Aroclor-1248-3       | (3) | 5.21       | 5.22      | 5.12      | 5.32 | 0.01       |
| Aroclor-1248-4       | (4) | 5.38       | 5.39      | 5.29      | 5.49 | 0.01       |
| Aroclor-1248-5       | (5) | 5.78       | 5.78      | 5.68      | 5.88 | 0.00       |
| Tetrachloro-m-xylene |     | 3.84       | 3.84      | 3.74      | 3.94 | 0.00       |
| Decachlorobiphenyl   |     | 8.90       | 8.91      | 8.81      | 9.01 | 0.01       |