

CALIBRATION VERIFICATION SUMMARY

Contract: SPEC01

Lab Code: CHEM Case No.: Q1322 SAS No.: Q1322 SDG NO.: Q1322

Continuing Calib Date: 02/07/2025 Initial Calibration Date(s): 01/28/2025 01/28/2025

Continuing Calib Time: 20:01 Initial Calibration Time(s): 09:37 17:30

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	5.69	5.69	5.59	5.79	0.00
Aroclor-1242-2	(2)	5.71	5.71	5.61	5.81	0.00
Aroclor-1242-3	(3)	5.77	5.77	5.67	5.87	0.00
Aroclor-1242-4	(4)	5.87	5.87	5.77	5.97	0.00
Aroclor-1242-5	(5)	6.60	6.60	6.50	6.70	0.00
Tetrachloro-m-xylene		4.53	4.54	4.44	4.64	0.01
Decachlorobiphenyl		10.26	10.28	10.18	10.38	0.02

CALIBRATION VERIFICATION SUMMARY

Contract: SPEC01

Lab Code: CHEM Case No.: Q1322 SAS No.: Q1322 SDG NO.: Q1322

Continuing Calib Date: 02/07/2025 Initial Calibration Date(s): 01/28/2025 01/28/2025

Continuing Calib Time: 20:01 Initial Calibration Time(s): 09:37 17:30

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.93	4.94	4.84	5.04	0.01
Aroclor-1242-2	(2)	4.95	4.95	4.85	5.05	0.00
Aroclor-1242-3	(3)	5.12	5.13	5.03	5.23	0.01
Aroclor-1242-4	(4)	5.21	5.22	5.12	5.32	0.01
Aroclor-1242-5	(5)	5.73	5.74	5.64	5.84	0.01
Tetrachloro-m-xylene		3.84	3.84	3.74	3.94	0.00
Decachlorobiphenyl		8.90	8.91	8.81	9.01	0.01