

CALIBRATION VERIFICATION SUMMARY

Contract: SPEC01

Lab Code: CHEM Case No.: Q1322 SAS No.: Q1322 SDG NO.: Q1322

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL02 Date Analyzed: 02/07/2025

Lab Sample No.: AR1660CCC500 Data File : PP069617.D Time Analyzed: 19:45

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.687	5.592	5.792	546.340	500.000	9.3
Aroclor-1016-2	5.709	5.614	5.814	550.630	500.000	10.1
Aroclor-1016-3	5.772	5.676	5.876	537.240	500.000	7.4
Aroclor-1016-4	5.869	5.774	5.974	546.830	500.000	9.4
Aroclor-1016-5	6.161	6.067	6.267	546.940	500.000	9.4
Aroclor-1260-1	7.280	7.187	7.387	554.520	500.000	10.9
Aroclor-1260-2	7.534	7.441	7.641	522.170	500.000	4.4
Aroclor-1260-3	7.893	7.800	8.000	534.210	500.000	6.8
Aroclor-1260-4	8.117	8.025	8.225	514.170	500.000	2.8
Aroclor-1260-5	8.438	8.347	8.547	516.030	500.000	3.2
Decachlorobiphenyl	10.264	10.179	10.379	54.890	50.000	9.8
Tetrachloro-m-xylene	4.535	4.438	4.638	56.360	50.000	12.7

CALIBRATION VERIFICATION SUMMARY

Contract: SPEC01

Lab Code: CHEM Case No.: Q1322 SAS No.: Q1322 SDG NO.: Q1322

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL02 Date Analyzed: 02/07/2025

Lab Sample No.: AR1660CCC500 Data File : PP069617.D Time Analyzed: 19:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.928	4.834	5.034	549.010	500.000	9.8
Aroclor-1016-2	4.947	4.853	5.053	547.990	500.000	9.6
Aroclor-1016-3	5.124	5.030	5.230	550.750	500.000	10.2
Aroclor-1016-4	5.166	5.072	5.272	533.160	500.000	6.6
Aroclor-1016-5	5.381	5.288	5.488	547.650	500.000	9.5
Aroclor-1260-1	6.418	6.326	6.526	575.420	500.000	15.1
Aroclor-1260-2	6.607	6.515	6.715	571.790	500.000	14.4
Aroclor-1260-3	6.761	6.670	6.870	496.600	500.000	-0.7
Aroclor-1260-4	7.234	7.142	7.342	559.340	500.000	11.9
Aroclor-1260-5	7.475	7.384	7.584	559.740	500.000	11.9
Decachlorobiphenyl	8.898	8.811	9.011	53.340	50.000	6.7
Tetrachloro-m-xylene	3.837	3.740	3.940	55.200	50.000	10.4