

Surrogate Summary

SDG No.: Q1322

Client: Spectra East Inc.

Analytical Method: 608.3 Pest

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL093725.D	PIBLK-PL093725.D	Decachlorobiphenyl	1	20	22.1	111		43	140
		Tetrachloro-m-xylene	1	20	20.8	104		77	126
		Decachlorobiphenyl	2	20	21.9	109		43	140
		Tetrachloro-m-xylene	2	20	20.5	103		77	126
I.BLK-PL094108.D	PIBLK-PL094108.D	Decachlorobiphenyl	1	20	19.2	96		43	140
		Tetrachloro-m-xylene	1	20	18.7	94		77	126
		Decachlorobiphenyl	2	20	18.9	94		43	140
		Tetrachloro-m-xylene	2	20	18.2	91		77	126
PB166620BL	PB166620BL	Tetrachloro-m-xylene	1	20	25.1	126		60	140
		Decachlorobiphenyl	1	20	24.9	124		60	140
		Tetrachloro-m-xylene	2	20	24.6	123		60	140
		Decachlorobiphenyl	2	20	25.1	126		60	140
PB166620BS	PB166620BS	Tetrachloro-m-xylene	1	20	21.6	108		60	140
		Decachlorobiphenyl	1	20	21.0	105		60	140
		Tetrachloro-m-xylene	2	20	20.7	104		60	140
		Decachlorobiphenyl	2	20	20.8	104		60	140
PB166620BSD	PB166620BSD	Tetrachloro-m-xylene	1	20	20.6	103		60	140
		Decachlorobiphenyl	1	20	19.8	99		60	140
		Tetrachloro-m-xylene	2	20	19.6	98		60	140
		Decachlorobiphenyl	2	20	20.1	100		60	140
Q1322-01	MANHOLE	Tetrachloro-m-xylene	1	20	17.1	85		60	140
		Decachlorobiphenyl	1	20	12.9	65		60	140
		Tetrachloro-m-xylene	2	20	18.5	92		60	140
		Decachlorobiphenyl	2	20	12.7	64		60	140
I.BLK-PL094119.D	PIBLK-PL094119.D	Decachlorobiphenyl	1	20	19.2	96		43	140
		Tetrachloro-m-xylene	1	20	18.3	92		77	126
		Decachlorobiphenyl	2	20	19.6	98		43	140
		Tetrachloro-m-xylene	2	20	18.2	91		77	126