

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020625\
 Data File : BF141479.D
 Acq On : 06 Feb 2025 14:14
 Operator : RC/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC080

Quant Time: Feb 06 16:01:02 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020625.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 06 15:49:10 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.792	152	90415	20.000	ng	0.00
21) Naphthalene-d8	8.075	136	341252	20.000	ng	0.00
39) Acenaphthene-d10	9.833	164	186230	20.000	ng	0.00
64) Phenanthrene-d10	11.316	188	312902	20.000	ng	0.00
76) Chrysene-d12	13.962	240	184347	20.000	ng	0.00
86) Perylene-d12	15.415	264	170605	20.000	ng	0.00

System Monitoring Compounds						
5) 2-Fluorophenol	5.410	112	878184	152.272	ng	0.00
7) Phenol-d6	6.439	99	1112281	152.591	ng	0.00
23) Nitrobenzene-d5	7.363	82	1048314	160.228	ng	0.00
42) 2,4,6-Tribromophenol	10.627	330	291947	156.058	ng	0.00
45) 2-Fluorobiphenyl	9.151	172	1789795	148.066	ng	0.00
79) Terphenyl-d14	12.904	244	1781602	163.152	ng	0.00

Target Compounds					Qvalue	
2) 1,4-Dioxane	2.481	88	181328	78.119	ng	99
3) Pyridine	3.216	79	435462	78.838	ng	99
4) n-Nitrosodimethylamine	3.187	42	301403	82.409	ng	97
6) Aniline	6.463	93	498556	72.913	ng	94
8) 2-Chlorophenol	6.586	128	470975	75.577	ng	99
10) Phenol	6.457	94	567332	74.779	ng	99
11) bis(2-Chloroethyl)ether	6.534	93	454141	79.695	ng	99
12) 1,3-Dichlorobenzene	6.734	146	494180	75.115	ng	99
13) 1,4-Dichlorobenzene	6.810	146	498585	74.175	ng	99
14) 1,2-Dichlorobenzene	6.963	146	464559	74.465	ng	99
15) Benzyl Alcohol	6.939	79	420673	75.258	ng	100
16) 2,2'-oxybis(1-Chloropr...	7.069	45	708318	72.613	ng	77
17) 2-Methylphenol	7.057	107	370173	75.912	ng	97
18) Hexachloroethane	7.304	117	190062	76.402	ng	97
19) n-Nitroso-di-n-propyla...	7.216	70	328097	75.249	ng	98
20) 3+4-Methylphenols	7.216	107	450046	72.613	ng	# 88
22) Acetophenone	7.210	105	652334	76.846	ng	98
24) Nitrobenzene	7.381	77	513249	80.448	ng	98
25) Isophorone	7.622	82	838076	80.337	ng	99
26) 2-Nitrophenol	7.692	139	261386	82.350	ng	98
27) 2,4-Dimethylphenol	7.733	122	300036	80.915	ng	98
28) bis(2-Chloroethoxy)met...	7.828	93	529442	79.203	ng	99
29) 2,4-Dichlorophenol	7.939	162	399027	79.645	ng	98
30) 1,2,4-Trichlorobenzene	8.016	180	428656	78.568	ng	99
31) Naphthalene	8.098	128	1381453	77.769	ng	100
32) Benzoic acid	7.892	122	335418	87.086	ng	98
33) 4-Chloroaniline	8.157	127	504834	81.400	ng	98
34) Hexachlorobutadiene	8.210	225	257532	78.628	ng	100
35) Caprolactam	8.545	113	125619	82.350	ng	97
36) 4-Chloro-3-methylphenol	8.639	107	441925	79.630	ng	98
37) 2-Methylnaphthalene	8.786	142	889774	76.975	ng	100
38) 1-Methylnaphthalene	8.886	142	861860	76.983	ng	99
40) 1,2,4,5-Tetrachloroben...	8.951	216	418579	77.689	ng	99
41) Hexachlorocyclopentadiene	8.939	237	159430	88.965	ng	98
43) 2,4,6-Trichlorophenol	9.069	196	277156	79.591	ng	97
44) 2,4,5-Trichlorophenol	9.116	196	298856	79.189	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020625\
 Data File : BF141479.D
 Acq On : 06 Feb 2025 14:14
 Operator : RC/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC080

Quant Time: Feb 06 16:01:02 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020625.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 06 15:49:10 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) 1,1'-Biphenyl	9.257	154	1100720	76.237	ng	99
47) 2-Chloronaphthalene	9.280	162	821334	76.929	ng	99
48) 2-Nitroaniline	9.380	65	286532	79.969	ng	97
49) Acenaphthylene	9.692	152	1211515	76.291	ng	99
50) Dimethylphthalate	9.563	163	1002121	79.265	ng	99
51) 2,6-Dinitrotoluene	9.622	165	215853	78.101	ng	94
52) Acenaphthene	9.869	154	819818	76.791	ng	99
53) 3-Nitroaniline	9.792	138	235511	79.173	ng	97
54) 2,4-Dinitrophenol	9.904	184	147768	89.960	ng	95
55) Dibenzofuran	10.039	168	1174157	74.928	ng	100
56) 4-Nitrophenol	9.969	139	187131	84.744	ng	97
57) 2,4-Dinitrotoluene	10.027	165	284954	77.449	ng	95
58) Fluorene	10.380	166	909704	75.080	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.163	232	253454	78.006	ng	98
60) Diethylphthalate	10.257	149	945293	75.995	ng	100
61) 4-Chlorophenyl-phenyle...	10.374	204	435522	74.716	ng	99
62) 4-Nitroaniline	10.416	138	241257	79.873	ng	99
63) Azobenzene	10.533	77	963689	77.931	ng	99
65) 4,6-Dinitro-2-methylph...	10.439	198	185427	85.309	ng	99
66) n-Nitrosodiphenylamine	10.498	169	785313	78.404	ng	99
67) 4-Bromophenyl-phenylether	10.863	248	276228	78.988	ng	96
68) Hexachlorobenzene	10.927	284	285111	79.174	ng	97
69) Atrazine	11.021	200	219923	73.188	ng	100
70) Pentachlorophenol	11.127	266	200556	87.100	ng	99
71) Phenanthrene	11.345	178	1312474	76.201	ng	100
72) Anthracene	11.398	178	1322462	76.380	ng	99
73) Carbazole	11.551	167	1168253	74.989	ng	99
74) Di-n-butylphthalate	11.880	149	1382706	76.865	ng	99
75) Fluoranthene	12.533	202	1333237	73.011	ng	100
77) Benzidine	12.657	184	354555	87.207	ng	99
78) Pyrene	12.763	202	1319554	84.086	ng	100
80) Butylbenzylphthalate	13.380	149	459803	84.592	ng	100
81) Benzo(a)anthracene	13.951	228	943291	76.977	ng	100
82) 3,3'-Dichlorobenzidine	13.915	252	263773	75.143	ng	100
83) Chrysene	13.992	228	906415	80.108	ng	100
84) Bis(2-ethylhexyl)phtha...	13.939	149	517884	84.477	ng	99
85) Di-n-octyl phthalate	14.562	149	803601	91.886	ng	100
87) Indeno(1,2,3-cd)pyrene	16.874	276	961023	91.166	ng	100
88) Benzo(b)fluoranthene	15.004	252	865205	75.529	ng	100
89) Benzo(k)fluoranthene	15.033	252	776323	79.364	ng	99
90) Benzo(a)pyrene	15.362	252	742932	80.150	ng	99
91) Dibenzo(a,h)anthracene	16.892	278	773684	90.578	ng	99
92) Benzo(g,h,i)perylene	17.315	276	831418	93.015	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Feb 06 16:01:02 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020625.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 06 15:49:10 2025
Response via : Initial Calibration

