

Surrogate Summary

SW-846

SDG No.: Q1322

Client: Spectra East Inc.

Analytical Method: 625.1

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB166619BL	PB166619BL	Nitrobenzene-d5	100	90.2	90		60	140
		2-Fluorobiphenyl	100	90.0	90		60	140
		Terphenyl-d14	100	81.1	81		60	140
PB166619BS	PB166619BS	Nitrobenzene-d5	100	95.0	95		60	140
		2-Fluorobiphenyl	100	96.3	96		60	140
		Terphenyl-d14	100	102	102		60	140
PB166619BSD	PB166619BSD	Nitrobenzene-d5	100	94.9	95		60	140
		2-Fluorobiphenyl	100	95.9	96		60	140
		Terphenyl-d14	100	94.7	95		60	140
Q1322-01	MANHOLE	Nitrobenzene-d5	100	76.4	76		60	140
		2-Fluorobiphenyl	100	78.4	78		60	140
		Terphenyl-d14	100	89.2	89		60	140