

Report of Analysis

Client:	Spectra East Inc.		Date Collected:	
Project:	Outfall 001 - Orangetown Dis Permit 2025		Date Received:	
Client Sample ID:	VX0207WBSD01		SDG No.:	Q1322
Lab Sample ID:	VX0207WBSD01		Matrix:	Water
Analytical Method:	E624.1		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044858.D	1		02/07/25 11:01	VX020725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-09-2	Methylene Chloride	21.4		1.20	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	20.8		0.95	5.00	ug/L
67-66-3	Chloroform	21.0		0.72	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.5		0.79	5.00	ug/L
79-01-6	Trichloroethene	20.2		0.77	5.00	ug/L
108-88-3	Toluene	20.7		0.72	5.00	ug/L
127-18-4	Tetrachloroethene	20.0		0.94	5.00	ug/L
100-41-4	Ethyl Benzene	20.8		0.73	5.00	ug/L
1330-20-7	Total Xylenes	64.0		2.52	15.0	ug/L
541-73-1	1,3-Dichlorobenzene	21.7		0.88	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	29.4		91 - 110	98%	SPK: 30
2037-26-5	Toluene-d8	29.3		91 - 112	98%	SPK: 30
460-00-4	4-Bromofluorobenzene	29.1		63 - 112	97%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	19400	4.891			
540-36-3	1,4-Difluorobenzene	109000	6.757			
3114-55-4	Chlorobenzene-d5	100000	10.049			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products