

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1322

Client: Spectra East Inc.

Analytical Method: SW624.1

Datafile : VX044857.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0207WBS01	Methylene Chloride	20	21.3	ug/L	106			60	140	
	trans-1,2-Dichloroethene	20	21.5	ug/L	108			70	130	
	Chloroform	20	21.0	ug/L	105			70	135	
	1,1,1-Trichloroethane	20	19.2	ug/L	96			70	130	
	Trichloroethene	20	19.7	ug/L	99			65	135	
	Toluene	20	20.7	ug/L	104			70	130	
	Tetrachloroethene	20	20.9	ug/L	104			70	130	
	Ethyl Benzene	20	20.8	ug/L	104			60	140	
	m/p-Xylenes	40	43.0	ug/L	108			87	111	
	o-Xylene	20	21.5	ug/L	108			87	111	
	1,3-Dichlorobenzene	20	22.1	ug/L	111			70	130	