

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1322

Client: Spectra East Inc.

Analytical Method: SW624.1

Datafile : VX044858.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0207WBSD01	Methylene Chloride	20	21.4	ug/L	107	1		60	140	20
	trans-1,2-Dichloroethene	20	20.8	ug/L	104	4		70	130	20
	Chloroform	20	21.0	ug/L	105	0		70	135	20
	1,1,1-Trichloroethane	20	19.5	ug/L	98	2		70	130	20
	Trichloroethene	20	20.2	ug/L	101	2		65	135	20
	Toluene	20	20.7	ug/L	104	0		70	130	20
	Tetrachloroethene	20	20.0	ug/L	100	4		70	130	20
	Ethyl Benzene	20	20.8	ug/L	104	0		60	140	20
	m/p-Xylenes	40	42.8	ug/L	107	1		87	111	20
	o-Xylene	20	21.2	ug/L	106	2		87	111	20
	1,3-Dichlorobenzene	20	21.7	ug/L	109	2		70	130	20