

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: SPEC01
 Lab Code: CHEM Case No.: Q1322 SAS No.: Q1322 SDG NO.: Q1322
 Lab File ID: VX044856.D Date Analyzed: 02/07/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:01
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	21982	4.89	127777	6.75	110326	10.05
UPPER LIMIT	43964	5.391	255554	7.251	220652	10.549
LOWER LIMIT	10991	4.391	63888.5	6.251	55163	9.549
EPA SAMPLE NO.						
MANHOLE	20788	4.90	113802	6.76	102149	10.05
VX0207WBL01	20510	4.90	114099	6.76	99648	10.05
VX0207WBS01	21269	4.89	121245	6.75	109254	10.05
VX0207WBSD01	19387	4.89	108776	6.76	100274	10.05

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.