



SAMPLE DATA

Report of Analysis

Client:	VERINA CONSULTING GROUP, LLC	Date Collected:	02/06/25 13:20
Project:	Rotor Clip NJ WTD - 2025	Date Received:	02/06/25
Client Sample ID:	WATER TREATMENT DISCHARGE	SDG No.:	Q1324
Lab Sample ID:	Q1324-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Ammonia as N	2.10	OR	1	0.045	0.10	mg/L	02/07/25 09:10	02/07/25 12:33	SM 4500-NH3 B plus G-11
Residual Chlorine	1.12	H	1	0.016	0.10	mg/L		02/07/25 10:42	SM 4500-Cl G-11

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	VERINA CONSULTING GROUP, LLC	Date Collected:	02/06/25 13:20
Project:	Rotor Clip NJ WTD - 2025	Date Received:	02/06/25
Client Sample ID:	WATER TREATMENT DISCHARGEDL	SDG No.:	Q1324
Lab Sample ID:	Q1324-01DL	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Ammonia as N	2.40	D	2	0.090	0.20	mg/L	02/07/25 09:10	02/07/25 13:34	SM 4500-NH3 B plus G-11

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits



QC RESULT SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1324

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB134619

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Residual Chlorine	mg/L	0.368	0.4	92	90-110	02/07/2025
Sample ID: CCV1 Residual Chlorine	mg/L	0.378	0.4	94	90-110	02/07/2025
Sample ID: CCV2 Residual Chlorine	mg/L	0.378	0.4	94	90-110	02/07/2025

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1324

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB134623

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	02/07/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.99	1	99	90-110	02/07/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	02/07/2025
Sample ID: CCV3 Ammonia as N	mg/L	1.1	1	110	90-110	02/07/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.97	1	97	90-110	02/07/2025



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Blank Summary

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1324

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB134619

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Residual Chlorine	mg/L	< 0.0500	0.0500	U	0.016	0.1	02/07/2025
Sample ID: CCB1 Residual Chlorine	mg/L	< 0.0500	0.0500	U	0.016	0.1	02/07/2025
Sample ID: CCB2 Residual Chlorine	mg/L	< 0.0500	0.0500	U	0.016	0.1	02/07/2025

Initial and Continuing Calibration Blank Summary

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1324

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB134623

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	02/07/2025
Sample ID: CCB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	02/07/2025
Sample ID: CCB2 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	02/07/2025
Sample ID: CCB3 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	02/07/2025
Sample ID: CCB4 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	02/07/2025

Preparation Blank Summary

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1324

Project: Rotor Clip NJ WTD - 2025

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB134619BL							
Residual Chlorine	mg/L	< 0.0500	0.0500	U	0.016	0.1	02/07/2025
Sample ID: PB166612BL							
Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	02/07/2025

Matrix Spike Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q1324
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q1324-01
Client ID:	WATER TREATMENT DISCHARGEMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Residual Chlorine	mg/L	71-148	1.51		1.12		0.4	1	98		02/07/2025

Matrix Spike Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q1324
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q1324-01
Client ID:	WATER TREATMENT DISCHARGEMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Residual Chlorine	mg/L	71-148	1.46		1.12		0.4	1	85		02/07/2025

Matrix Spike Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q1324
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q1325-01
Client ID:	DSN002MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/L	75-125	7.70	OR	6.90	OR	1	1	80		02/07/2025

Matrix Spike Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q1324
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q1325-01
Client ID:	DSN002MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/L	75-125	8.20	OR	6.90	OR	1	1	130	*	02/07/2025

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q1324
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q1324-01
Client ID:	WATER TREATMENT DISCHARGEDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Residual Chlorine	mg/L	+/-20	1.12		1.17		1	4.37		02/07/2025

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q1324
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q1324-01
Client ID:	WATER TREATMENT DISCHARGEMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Residual Chlorine	mg/L	+/-20	1.51		1.46		1	3.37		02/07/2025

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q1324
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q1325-01
Client ID:	DSN002DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	6.90	OR	6.90	OR	1	0		02/07/2025

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q1324
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q1325-01
Client ID:	DSN002MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	7.70	OR	8.20	OR	1	6		02/07/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q1324
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB134619

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134619BS							
Residual Chlorine	mg/L	0.4	0.39		97	1	90-110	02/07/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q1324
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB134623

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB166612BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	02/07/2025

Instrument ID: SPECTROPHOTOMETER-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB134619

Review By	Niha	Review On	2/7/2025 12:55:11 PM
Supervise By	Iwona	Supervise On	2/7/2025 2:49:27 PM
SubDirectory	LB134619	Test	Residual Chlorine
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	WP111848,WP111847,WP111842,WP111843,WP111844,WP111841,WP111845,WP111846,W3147		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	CAL1	CAL1	CAL	02/07/25 10:00		Niha	OK
2	CAL2	CAL2	CAL	02/07/25 10:03		Niha	OK
3	CAL3	CAL3	CAL	02/07/25 10:06		Niha	OK
4	CAL4	CAL4	CAL	02/07/25 10:09		Niha	OK
5	CAL5	CAL5	CAL	02/07/25 10:12		Niha	OK
6	CAL6	CAL6	CAL	02/07/25 10:15		Niha	OK
7	ICV	ICV	ICV	02/07/25 10:18		Niha	OK
8	ICB	ICB	ICB	02/07/25 10:21		Niha	OK
9	CCV1	CCV1	CCV	02/07/25 10:24		Niha	OK
10	CCB1	CCB1	CCB	02/07/25 10:30		Niha	OK
11	LB134619BL	LB134619BL	MB	02/07/25 10:33		Niha	OK
12	LB134619BS	LB134619BS	LCS	02/07/25 10:36		Niha	OK
13	Q1168-09	MDL-WATER-03-QT1	SAM	02/07/25 10:39		Niha	OK
14	Q1324-01	WATER TREATMENT	SAM	02/07/25 10:42		Niha	OK
15	Q1324-01DUP	WATER TREATMENT	DUP	02/07/25 10:45		Niha	OK
16	Q1324-01MS	WATER TREATMENT	MS	02/07/25 10:48		Niha	OK
17	Q1324-01MSD	WATER TREATMENT	MSD	02/07/25 10:51		Niha	OK
18	CCV2	CCV2	CCV	02/07/25 10:54		Niha	OK

Instrument ID: SPECTROPHOTOMETER-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB134619

Review By	Niha	Review On	2/7/2025 12:55:11 PM
Supervise By	Iwona	Supervise On	2/7/2025 2:49:27 PM
SubDirectory	LB134619	Test	Residual Chlorine
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	WP111848,WP111847,WP111842,WP111843,WP111844,WP111841,WP111845,WP111846,W3147		

19	CCB2	CCB2	CCB	02/07/25 10:57		Niha	OK
----	------	------	-----	----------------	--	------	----

A
B
C
D
E
F

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB134623

Review By	rubina	Review On	2/10/2025 8:55:56 AM
Supervise By	Iwona	Supervise On	2/10/2025 9:22:59 AM
SubDirectory	LB134623	Test	Ammonia
STD. NAME	STD REF.#		
ICAL Standard	WP111829		
ICV Standard	WP111831		
CCV Standard	WP111830		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP111420		
Chk Standard	WP110416,WP111745,WP111385,WP111660,WP111832		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	0.0PPM	0.0PPM	CAL1	02/07/25 11:18		rubina	OK
2	0.1PPM	0.1PPM	CAL2	02/07/25 11:18		rubina	OK
3	0.2PPM	0.2PPM	CAL3	02/07/25 11:18		rubina	OK
4	0.4PPM	0.4PPM	CAL4	02/07/25 11:18		rubina	OK
5	1.0PPM	1.0PPM	CAL5	02/07/25 11:18		rubina	OK
6	1.3PPM	1.3PPM	CAL6	02/07/25 11:18		rubina	OK
7	2.0PPM	2.0PPM	CAL7	02/07/25 11:18		rubina	OK
8	ICV1	ICV1	ICV	02/07/25 12:22		rubina	OK
9	ICB1	ICB1	ICB	02/07/25 12:22		rubina	OK
10	CCV1	CCV1	CCV	02/07/25 12:22		rubina	OK
11	CCB1	CCB1	CCB	02/07/25 12:22		rubina	OK
12	RL	RL	SAM	02/07/25 12:22		rubina	OK
13	PB166612BL	PB166612BL	MB	02/07/25 12:22		rubina	OK
14	PB166612BS	PB166612BS	LCS	02/07/25 12:33		rubina	OK
15	Q1316-01	001-WILLETS-PT-BL	SAM	02/07/25 12:33	High	rubina	Dilution
16	Q1316-02	002-35TH-AVE(FEB)	SAM	02/07/25 12:33	High	rubina	Dilution
17	Q1322-04	MANHOLE	SAM	02/07/25 12:33		rubina	OK
18	Q1324-01	WATER TREATMENT	SAM	02/07/25 12:33	High	rubina	Dilution

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB134623

Review By	rubina	Review On	2/10/2025 8:55:56 AM
Supervise By	Iwona	Supervise On	2/10/2025 9:22:59 AM
SubDirectory	LB134623	Test	Ammonia
STD. NAME	STD REF.#		
ICAL Standard	WP111829		
ICV Standard	WP111831		
CCV Standard	WP111830		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP111420		
Chk Standard	WP110416,WP111745,WP111385,WP111660,WP111832		

19	Q1325-01	DSN002	SAM	02/07/25 12:33	High	rubina	Dilution
20	Q1325-01DUP	DSN002DUP	DUP	02/07/25 12:33		rubina	OK
21	CCV2	CCV2	CCV	02/07/25 12:44		rubina	OK
22	CCB2	CCB2	CCB	02/07/25 12:44		rubina	OK
23	Q1325-01MS	DSN002MS	MS	02/07/25 12:44		rubina	OK
24	Q1325-01MSD	DSN002MSD	MSD	02/07/25 12:44		rubina	OK
25	Q1325-03	DSN001	SAM	02/07/25 12:44		rubina	OK
26	Q1325-05	DSN003	SAM	02/07/25 12:44	High	rubina	Dilution
27	PB166614BL	PB166614BL	MB	02/07/25 12:44		rubina	OK
28	PB166614BS	PB166614BS	LCS	02/07/25 12:44		rubina	OK
29	CCV3	CCV3	CCV	02/07/25 12:52		rubina	OK
30	CCB3	CCB3	CCB	02/07/25 12:52		rubina	OK
31	Q1168-09	MDL-WATER-03-QT1	SAM	02/07/25 13:25		rubina	OK
32	Q1168-03	MDL-SOIL-03-QT1-20	SAM	02/07/25 13:25		rubina	OK
33	Q1316-01DL	001-WILLETTS-PT-BLV	SAM	02/07/25 13:34	Report 10X	rubina	Confirms
34	Q1316-02DL	002-35TH-AVE(FEB)D	SAM	02/07/25 13:34	Report 10X	rubina	Confirms
35	Q1324-01DL	WATER TREATMENT	SAM	02/07/25 13:34	Report 2X	rubina	Confirms
36	Q1325-01DL	DSN002DL	SAM	02/07/25 13:35	Report 10X	rubina	Confirms
37	Q1325-05DL	DSN003DL	SAM	02/07/25 13:35	Report 10X	rubina	Confirms
38	CCV4	CCV4	CCV	02/07/25 13:35		rubina	OK

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB134623

Review By	rubina	Review On	2/10/2025 8:55:56 AM
Supervise By	Iwona	Supervise On	2/10/2025 9:22:59 AM
SubDirectory	LB134623	Test	Ammonia
STD. NAME	STD REF.#		
ICAL Standard	WP111829		
ICV Standard	WP111831		
CCV Standard	WP111830		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP111420		
Chk Standard	WP110416,WP111745,WP111385,WP111660,WP111832		

39	CCB4	CCB4	CCB	02/07/25 13:35		rubina	OK
----	------	------	-----	----------------	--	--------	----

A
B
C
D
E
F

LAB CHRONICLE

OrderID:	Q1324	OrderDate:	2/6/2025 1:55:00 PM
Client:	VERINA CONSULTING GROUP, LLC	Project:	Rotor Clip NJ WTD - 2025
Contact:	Michael Valenzi	Location:	D11

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1324-01	WATER TREATMENT DISCHARGE	WATER			02/06/25 13:20	02/07/25		02/06/25
			Ammonia	SM4500-NH3			02/07/25 12:33	
			Residual Chlorine	SM4500 Cl G			02/07/25 10:42	
Q1324-01DL	WATER TREATMENT DISCHARGEDL	WATER			02/06/25 13:20	02/07/25		02/06/25
			Ammonia	SM4500-NH3			02/07/25 13:34	

SOP ID : MSM4500-NH3 B,G-Ammonia-17

SDG No : N/A

Start Digest Date: 02/07/2025 Time : 09:10 Temp : 150 °C

Matrix : WATER

End Digest Date: 02/07/2025 Time : 10:10 Temp : 160 °C

Pipette ID : WC

if batch 02/07/2025 10:30 150°C } RM
02/07/2025 11:30 160°C }

Balance ID : N/A

Hood ID : HOOD#2

Digestion tube ID : M5595

Block Thermometer ID : WC CYANIDE

Block ID : WC-DIST-BLOCK-1

Filter paper ID : N/A

Prep Technician Signature: RM

Weigh By : N/A

pH Meter ID : N/A

Supervisor Signature: 12

Standard Name	MLS USED	STD REF. # FROM LOG
LCSW	1.0ML	WP111420
MS/MSD SPIKE SOL.	1.0ML	WP111419
PBW	50.0ML	W3112
RL CHECK	0.1ML	WP111419
MDL	0.8ML	WP111832

Chemical Used	ML/SAMPLE USED	Lot Number
BORATE BUFFER	2.5ML	WP111325
NAOH 6N	0.5-2.0ML	WP111318
H2SO4 0.04N	5.0ML	WP110335
pH strip-Ammonia	N/A	W3133
KI-starch paper	N/A	W3155
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

ALL GLASSWEAR ARE STEAMED OUT AND THERE WERE NO TRACE OF AMMONIA USING NESLER REAGENT
WP111604,

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
02/07/2025 11:45	RM CWL	RM CWL
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Vol (ml)	Final Vol (ml)	pH	Sulfide	Oxidizing	Nitrate/ Nitrite	Comment	Prep Pos
PB166612BL	PBW612	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
PB166612BS	LCS612	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q1168-09	MDL-WATER-03-QT1-2025	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q1316-01	001-WILLETS-PT-BLVD(FEB)	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q1316-02	002-35TH-AVE(FEB)	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q1322-04	MANHOLE	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q1324-01	WATER TREATMENT DISCHARGE	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q1325-01	DSN002	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q1325-01DUP	DSN002DUP	50	50	<2	N/A	Negative	N/A	PH AFTER ADDING DIST BUFFER>11	N/A
Q1325-01MS	DSN002MS	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q1325-01MSD	DSN002MSD	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q1325-03	DSN001	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q1325-05	DSN003	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A