



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1324

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB134619

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Residual Chlorine	mg/L	0.368	0.4	92	90-110	02/07/2025
Sample ID: CCV1 Residual Chlorine	mg/L	0.378	0.4	94	90-110	02/07/2025
Sample ID: CCV2 Residual Chlorine	mg/L	0.378	0.4	94	90-110	02/07/2025

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1324

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB134623

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	02/07/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.99	1	99	90-110	02/07/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	02/07/2025
Sample ID: CCV3 Ammonia as N	mg/L	1.1	1	110	90-110	02/07/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.97	1	97	90-110	02/07/2025