

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q1324
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1324 **SAS No.:** Q1324
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Chromium	528	520	102	90 - 110	P	02/12/2025	12:59	LB134704
	Copper	520	510	102	90 - 110	P	02/12/2025	12:59	LB134704
	Nickel	514	530	97	90 - 110	P	02/12/2025	12:59	LB134704
	Zinc	1010	1000	101	90 - 110	P	02/12/2025	12:59	LB134704

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q1324
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1324 **SAS No.:** Q1324
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Chromium	10.2	10.0	102	80 - 120	P	02/12/2025	13:08	LB134704
	Copper	21.9	20.0	110	80 - 120	P	02/12/2025	13:08	LB134704
	Nickel	39.3	40.0	98	80 - 120	P	02/12/2025	13:08	LB134704
	Zinc	42.2	40.0	106	80 - 120	P	02/12/2025	13:08	LB134704

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q1324

Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1324 **SAS No.:** Q1324

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Chromium	999	1000	100	90 - 110	P	02/12/2025	13:39	LB134704
	Copper	1260	1250	101	90 - 110	P	02/12/2025	13:39	LB134704
	Nickel	2500	2500	100	90 - 110	P	02/12/2025	13:39	LB134704
	Zinc	2470	2500	99	90 - 110	P	02/12/2025	13:39	LB134704
CCV02	Chromium	1030	1000	103	90 - 110	P	02/12/2025	14:29	LB134704
	Copper	1270	1250	102	90 - 110	P	02/12/2025	14:29	LB134704
	Nickel	2530	2500	101	90 - 110	P	02/12/2025	14:29	LB134704
	Zinc	2520	2500	101	90 - 110	P	02/12/2025	14:29	LB134704
CCV03	Chromium	1020	1000	102	90 - 110	P	02/12/2025	15:20	LB134704
	Copper	1250	1250	100	90 - 110	P	02/12/2025	15:20	LB134704
	Nickel	2490	2500	100	90 - 110	P	02/12/2025	15:20	LB134704
	Zinc	2510	2500	100	90 - 110	P	02/12/2025	15:20	LB134704
CCV04	Chromium	1020	1000	102	90 - 110	P	02/12/2025	16:13	LB134704
	Copper	1240	1250	99	90 - 110	P	02/12/2025	16:13	LB134704
	Nickel	2490	2500	100	90 - 110	P	02/12/2025	16:13	LB134704
	Zinc	2470	2500	99	90 - 110	P	02/12/2025	16:13	LB134704
CCV05	Chromium	1020	1000	102	90 - 110	P	02/12/2025	17:04	LB134704
	Copper	1250	1250	100	90 - 110	P	02/12/2025	17:04	LB134704
	Nickel	2490	2500	100	90 - 110	P	02/12/2025	17:04	LB134704
	Zinc	2480	2500	99	90 - 110	P	02/12/2025	17:04	LB134704
CCV06	Chromium	1020	1000	102	90 - 110	P	02/12/2025	17:55	LB134704
	Copper	1250	1250	100	90 - 110	P	02/12/2025	17:55	LB134704
	Nickel	2500	2500	100	90 - 110	P	02/12/2025	17:55	LB134704
	Zinc	2490	2500	99	90 - 110	P	02/12/2025	17:55	LB134704
CCV07	Chromium	1020	1000	102	90 - 110	P	02/12/2025	18:34	LB134704
	Copper	1270	1250	102	90 - 110	P	02/12/2025	18:34	LB134704
	Nickel	2510	2500	100	90 - 110	P	02/12/2025	18:34	LB134704
	Zinc	2470	2500	99	90 - 110	P	02/12/2025	18:34	LB134704