



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1324

Contract: VERI01

Lab Code: CHEM

Case No.: Q1324

SAS No.: Q1324

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Chromium	10.1	10.0	101	40 - 160	P	02/12/2025	13:17	LB134704
	Copper	22.2	20.0	111	40 - 160	P	02/12/2025	13:17	LB134704
	Nickel	40.4	40.0	101	40 - 160	P	02/12/2025	13:17	LB134704
	Zinc	43.0	40.0	108	40 - 160	P	02/12/2025	13:17	LB134704