

metals
- 5a -
MATRIX SPIKE SUMMARY

client: VERINA CONSULTING GROUP, LLC **level:** low **sdg no.:** Q1324
contract: VERI01 **lab code:** CHEM **case no.:** Q1324 **sas no.:** Q1324
matrix: Water **sample id:** Q1324-01 **client id:** WATER TREATMENT DISCHARGEMS
Percent Solids for Sample: NA **Spiked ID:** Q1324-01MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Chromium	ug/L	75 - 125	362		166		200	98		P
Copper	ug/L	75 - 125	149		13.1		150	90		P
Nickel	ug/L	75 - 125	259		6.78	J	250	101		P
Zinc	ug/L	75 - 125	269		163		100	106		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: VERINA CONSULTING GROUP, LLC **level:** low **sdg no.:** Q1324
contract: VERI01 **lab code:** CHEM **case no.:** Q1324 **sas no.:** Q1324
matrix: Water **sample id:** Q1324-01 **client id:** WATER TREATMENT DISCHARGEMSD
Percent Solids for Sample: NA **Spiked ID:** Q1324-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Chromium	ug/L	75 - 125	363		166		200	99		P
Copper	ug/L	75 - 125	148		13.1		150	90		P
Nickel	ug/L	75 - 125	259		6.78	J	250	101		P
Zinc	ug/L	75 - 125	267		163		100	105		P