

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1324

Contract: VERI01

Lab Code: CHEM

Case No.: Q1324

SAS No.: Q1324

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166635BS							
Chromium	ug/L	200	199		100	80 - 120	P
Copper	ug/L	150	146		97	80 - 120	P
Nickel	ug/L	250	236		94	80 - 120	P
Zinc	ug/L	100	100		100	80 - 120	P