



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q1325

Project: Linden Generating Station - LRSA

RunNo.: LB134609

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.500	0.5	100	90-110	02/06/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	02/06/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.501	0.5	100	90-110	02/06/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q1325

Project: Linden Generating Station - LRSA

RunNo.: LB134623

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	02/07/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.99	1	99	90-110	02/07/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	02/07/2025
Sample ID: CCV3 Ammonia as N	mg/L	1.1	1	110	90-110	02/07/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.97	1	97	90-110	02/07/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q1325

Project: Linden Generating Station - LRSA

RunNo.: LB134631

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	49.329	50	99	95-105	01/22/2025
Sample ID: CCV1 COD	mg/L	49.329	50	99	95-105	02/07/2025
Sample ID: CCV2 COD	mg/L	49.329	50	99	95-105	02/07/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q1325

Project: Linden Generating Station - LRSA

RunNo.: LB134656

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Phenolics	ICV1	mg/L	1	1	100	90-110	02/10/2025
Sample ID: Phenolics	CCV1	mg/L	0.99	1	99	90-110	02/10/2025
Sample ID: Phenolics	CCV2	mg/L	1	1	100	90-110	02/10/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q1325

Project: Linden Generating Station - LRSA

RunNo.: LB134710

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Field pH		pH	7.00	7	100	90-110	02/06/2025
Sample ID:	CCV1						
Field pH		pH	7.00	7	100	90-110	02/06/2025
Sample ID:	CCV2						
Field pH		pH	7.01	7	100	90-110	02/06/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q1325

Project: Linden Generating Station - LRSA

RunNo.: LB134710

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------