



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Blank Summary

Client: Parkway Generation Operating LLC

SDG No.: Q1325

Project: Linden Generating Station - LRSA

RunNo.: LB134609

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	02/06/2025
Sample ID: CCB1 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	02/06/2025
Sample ID: CCB2 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	02/06/2025

Initial and Continuing Calibration Blank Summary

Client: Parkway Generation Operating LLC

SDG No.: Q1325

Project: Linden Generating Station - LRSA

RunNo.: LB134623

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	02/07/2025
Sample ID: CCB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	02/07/2025
Sample ID: CCB2 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	02/07/2025
Sample ID: CCB3 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	02/07/2025
Sample ID: CCB4 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	02/07/2025

Initial and Continuing Calibration Blank Summary

Client: Parkway Generation Operating LLC

SDG No.: Q1325

Project: Linden Generating Station - LRSA

RunNo.: LB134631

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB COD	mg/L	< 5.0000	5.0000	U	2.35	10	01/22/2025
Sample ID: CCB1 COD	mg/L	< 5.0000	5.0000	U	2.35	10	02/07/2025
Sample ID: CCB2 COD	mg/L	< 5.0000	5.0000	U	2.35	10	02/07/2025

Initial and Continuing Calibration Blank Summary

Client: Parkway Generation Operating LLC

SDG No.: Q1325

Project: Linden Generating Station - LRSA

RunNo.: LB134656

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Phenolics	mg/L	< 0.0250	0.0250	U	0.023	0.05	02/10/2025
Sample ID: CCB1 Phenolics	mg/L	< 0.0250	0.0250	U	0.023	0.05	02/10/2025
Sample ID: CCB2 Phenolics	mg/L	< 0.0250	0.0250	U	0.023	0.05	02/10/2025

Initial and Continuing Calibration Blank Summary

Client: Parkway Generation Operating LLC

SDG No.: Q1325

Project: Linden Generating Station - LRSA

RunNo.: LB134656

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
---------	-------	--------	----------------------	--------------	-----	-----	------------------