

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1327 SAS No.: Q1327 SDG No.: Q1327

Instrument ID: BNA_N Calibration Date/Time: 02/13/2025 12:55

Lab File ID: BN036464.D Init. Calib. Date(s): 02/10/2025 02/10/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 12:25 16:00

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.615	0.600		-2.4	50.0
Fluoranthene-d10	1.112	1.045		-6.0	50.0
2-Fluorophenol	0.945	0.901		-4.7	50.0
Phenol-d6	1.109	1.071		-3.4	50.0
Nitrobenzene-d5	0.395	0.374		-5.3	50.0
2-Fluorobiphenyl	1.504	1.457		-3.1	50.0
2,4,6-Tribromophenol	0.198	0.166		-16.2	50.0
Terphenyl-d14	0.854	0.858		0.5	50.0
1,4-Dioxane	0.438	0.413		-5.7	50.0

All other compounds must meet a minimum RRF of 0.010.