

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021325\
 Data File : BN036459.D
 Acq On : 13 Feb 2025 09:48
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 13 10:19:40 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 01:17:14 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	103	0.00
2	1,4-Dioxane	0.400	0.393	1.8	103	0.00
3	n-Nitrosodimethylamine	0.400	0.392	2.0	101	0.00
4 S	2-Fluorophenol	0.400	0.384	4.0	100	0.00
5 S	Phenol-d6	0.400	0.393	1.8	109	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.399	0.3	110	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	109	0.00
8 S	Nitrobenzene-d5	0.400	0.377	5.8	112	-0.01
9	Naphthalene	0.400	0.387	3.3	109	-0.01
10	Hexachlorobutadiene	0.400	0.380	5.0	103	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.391	2.3	109	-0.01
12	2-Methylnaphthalene	0.400	0.386	3.5	108	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	118	0.00
14 S	2,4,6-Tribromophenol	0.400	0.323	19.3	102	0.00
15 S	2-Fluorobiphenyl	0.400	0.378	5.5	122	-0.01
16	Acenaphthylene	0.400	0.362	9.5	112	-0.01
17	Acenaphthene	0.400	0.370	7.5	112	0.00
18	Fluorene	0.400	0.369	7.8	110	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	109	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.326	18.5	101	0.00
21	4-Bromophenyl-phenylether	0.400	0.387	3.3	109	0.00
22	Hexachlorobenzene	0.400	0.391	2.3	110	-0.01
23	Atrazine	0.400	0.352	12.0	102	-0.01
24	Pentachlorophenol	0.400	0.337	15.8	105	-0.01
25	Phenanthrene	0.400	0.390	2.5	112	0.00
26	Anthracene	0.400	0.385	3.8	110	-0.01
27 SURR	Fluoranthene-d10	0.400	0.379	5.3	108	0.00
28	Fluoranthene	0.400	0.377	5.8	108	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	108	0.00
30	Pyrene	0.400	0.397	0.8	107	0.00
31 S	Terphenyl-d14	0.400	0.385	3.8	104	0.00
32	Benzo(a)anthracene	0.400	0.389	2.8	106	0.00
33	Chrysene	0.400	0.419	-4.7	118	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.340	15.0	97	0.00
35 I	Perylene-d12	0.400	0.400	0.0	98	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.361	9.8	90	-0.01
37	Benzo(b)fluoranthene	0.400	0.410	-2.5	105	0.00
38	Benzo(k)fluoranthene	0.400	0.416	-4.0	102	0.00
39 C	Benzo(a)pyrene	0.400	0.406	-1.5	104	0.00
40	Dibenzo(a,h)anthracene	0.400	0.357	10.8	90	-0.01
41	Benzo(g,h,i)perylene	0.400	0.356	11.0	87	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4